

## Certificado de Análise Analysis Certificate

**MS ANVISA** 80115310205  
**DESCRIÇÃO / DESCRIPTION** CREATININA WS R1 4X30ML + R2 2X15ML  
**LOTE / LOT:** 11015125119  
**DATA DE FABR./MANUF. DATE** 13/06/2025  
**DATA DE VALIDADE / EXPIRY DATE:** 17/04/2027

**DESCRIÇÃO / DESCRIPTION:** CREATININA WS R1 30ML (MIURA)  
**LOTE / LOT:** 11015225115  
**DATA DE VALIDADE / EXPIRY DATE:** 09/05/2027

**DESCRIÇÃO / DESCRIPTION:** CREATININA WS R2 15ML (MIURA)  
**LOTE / LOT:** 11015325113  
**DATA DE VALIDADE / EXPIRY DATE:** 17/04/2027

DESCRIÇÃO / DESCRIPTION	ESPECIFICAÇÃO / SPECIFICATION	RESULTADO / RESULTS
Odor / Odour R1	SEM ODOR CARACTERISTICO/NO CHARACTERISTIC ODOUR	Conforme/According
pH R1	13.0 (12.90-13.10)	12,97
Absorbancia / Absorbance R1	0.015 (0.000-0.030)	0
Absorbancia / Absorbance R1 + R2	0.150 (0.120-0.180)	0,154
Aporencia / Appearance R1	LIMPIDO, INCOLOR/CLEAR, COLOURLESS	Conforme/According
Aporencia / Appearance R2	LIMPIDO AMARELO/CLEAR, YELLOW	Conforme/According
Teste funcional / Functional test	DENTRO DOS VALORES DE REFERÊNCIA DOS CONTROLES/WITHIN THE REFERENCE VALUES OF THE CONTROLS	Conforme/According